

## **SPECIFICATIONS**

## Programmable DC Power Supply

MODEL: OPM-10030D



Parameter			Specifications		
Output rating(@0°C ~ 40°C)	Channel 1		0 to 100V / 0 to 30A		
Output fating(@0 C = 40 C)	Channel 2		0 to 100V / 0 to 30A		
Output WATT			6000W		
Programming Accuracy	Voltage		0.05%+33.3mV		
(@25℃ ±5℃)±(%of output + offset)	Current		0.2%+30.0mA		
Readback Accuracy	Voltage		0.05%+16.7mV		
25°C ±5°C)±(%of output + offset)		0.2%+15.0mA			
Diania and Naiss (2011, to 2011)	Voltage		≤ 0.01%mVrms		
Ripple and Noise(20Hz to 20MHz)	Current		≤ 7.5mArms		
1.0	Voltage		6.7mV		
Load Regulation	Current		3.0mA		
Line Regulation	Voltage		1.7mV		
	Current		3.0mA		
Resolution	Programming/Readback		≤0.83mV / ≤0.30mA		
	Display Meter		10mV / 1mA		
Temperature Coefficient ±(%of output + offset) Voltage			0.05%+10.0mV		
After a 30-minute warm-up	Current		0.2%+15.0mA		
Stability ±(%of output + offset)	Voltage		0.05%+3.3mV		
After a 1 hour warm-up	Current		0.2%+6.0mA		
	Sanoni		Less than 50//s for output to recover to within 15mV following a change in output current		
Transient Response Time			from full load to half load or vice versa		
	Τ	Rising time	≤ 7.5V/ms		
Voltage Programming Speed	No load	Falling time	≤ 3V/ms		
		Rising time	≤ 3.25V/ms		
	Half load		≤ 6V/ms		
	1		5% + 0.5V		
OVP and OCP Accuracy $\pm$ (%of output + offset			5% + 0.5A		
and Oor Accuracy ±(%or output + orrset	Activation Time		< 80ms when maximum output rating		
Tracking Accuracy	Activation	Time	0.1% + 10mV		
Tracking Accuracy	Power Switch ON/OFF		No overshoot, undershoot : ≤ -0.8V		
Output Voltage Overshoot & Undershoot	Voltage Output Setting		No overshoot, No undershoot		
Remote Interface		GPIB(IEEE-488.2) Option , RS232C Standard			
Programming Language			SCPI(Standard Commands for Programmable Instruments)		
r rogramming Language			Setting	28ms	
Command Processing Time(average)	Output Setting			32ms	
			Query Voltage & Current Setting	28ms	
			Voltage & Current Query	32ms	
			Voltage & Current Query		
	Measurement The Other		-	Present mode: 47ms Buffer mode: 32ms	
		Setting & Query	< 35ms		
State Storage Memory		Ten user-configurable(voltage,current,OVP & OCP level)stored states			
	Voltage Drop		Up to 1V per each lead		
Remote Sensing Capability	Load Regulation		Add 5 mV to spec for each 1-volt change in the + output lead due to load current changes.		
1	1 177 11				
	Load Voltage		Subtract voltage drop in load leads from specified output voltage atiing.		
Operation Temperature		0°C ~ 40°C for full rated output. At higher temperatures the output current is derated linearly to 50% at 55°C maximum temperature			
Cooling				temperature	
Cooling			Isolation AC FAN		
Output Terminal Isolated (maximum, from chassis ground)			$\pm 30$ V output is $\pm 60$ Vdc when connecting shorting conductors without insulation to the (+)output to the (+)sense and the (-)output and the (-)sense terminals		
Output Terminal Isolated (maximum, from chas				220V ± 10% 50~60Hz	
Output Terminal Isolated (maximum, from chas	Standard		220V ± 10% 50~60Hz		
	Standard		110V ± 10% 50~60Hz		
	Standard				
			110V ± 10% 50~60Hz		
AC Input Ratings			110V ± 10% 50~60Hz 115V ± 10% 50~60Hz		
	Option	nded	110V ± 10% 50~60Hz 115V ± 10% 50~60Hz 230V ± 10% 50~60Hz		
AC Input Ratings	Option Precision Recomme	nded	110V ± 10% 50~60Hz 115V ± 10% 50~60Hz 230V ± 10% 50~60Hz 6 month	n(D)	
AC Input Ratings  Calibration Interval	Option Precision Recomme	nded	110V ± 10% 50~60Hz 115V ± 10% 50~60Hz 230V ± 10% 50~60Hz 6 month 1 year	n(D)	
AC Input Ratings  Calibration Interval  Dimensions (19-inch * 18U Standard Rack Cas	Option Precision Recomme		110V ± 10% 50~60Hz 115V ± 10% 50~60Hz 230V ± 10% 50~60Hz 6 month 1 year 600mm(W) * 1000mm(H) * 750mm	n(D)	